

Annex of the certificate (Page 1/1)

Accreditation Scope

 Test TS EN ISO IEC 17025 AB-1535-T	LABİOTEK BİYOTEKNOLOJİ LABORATUVARLARI PROTEZ SANAYİ VE TİCARET LİMİTED ŞİRKETİ	
	Accreditation Nr: AB-1535-T Revision Nr: 00 Date: 29.01.2021	
As a Testing Laboratory		
Address: SÖĞÜTÖZÜ MAHALLESİ SÖĞÜTÖZÜ CADDESİ NO:43 (TOBB EKONOMİ VE TEKNOLOJİ ÜNİVERSİTESİ YERLEŞKESİ)	Phone : 03122924506 Fax : E-Mail : info@labiotech.com.tr Website : www.labiotech.com.tr	

Tested Materials / Products	Name of Test	Testing Method (National, International standards, in house methods)
INTRAMEDULLARY FIXATION DEVICES	TEST METHOD FOR STATIC FOUR-POINT BEND TEST	ASTM F 1264- ANNEX A.1
INTRAMEDULLARY FIXATION DEVICES	TEST METHOD FOR STATIC TORSIONAL TESTING OF INTRAMEDULLARY FIXATION DEVICES	ASTM F 1264- ANNEX A.2
INTRAMEDULLARY FIXATION DEVICES	TEST METHOD FOR BENDING FATIGUE TESTING OF INTRAMEDULLARY FIXATION DEVICES	ASTM F 1264- ANNEX A.3
INTRAMEDULLARY FIXATION DEVICES	TEST METHOD FOR BENDING FATIGUE TESTING OF IMFD LOCKING SCREWS	ASTM F 1264- ANNEX A.4
METALLIC BONE PLATES	STANDARD TEST METHOD FOR SINGLE CYCLE BEND TESTING OF METALLIC BONE PLATES	ASTM F 382 - ANNEX A.1
METALLIC BONE PLATES	STANDARD TEST METHOD FOR DETERMINING THE BENDING FATIGUE PROPERTIES OF METALLIC BONE PLATES	ASTM F 382 - ANNEX A.2
METALLIC MEDICAL BONE SCREWS	TEST METHOD FOR DETERMINING THE TORSIONAL PROPERTIES OF METALLIC BONE SCREWS	ASTM F 543 - ANNEX A.1
METALLIC MEDICAL BONE SCREWS	TEST METHOD FOR DETERMINING THE AXIAL PULLOUT STRENGTH OF MEDICAL BONE SCREWS	ASTM F 543 - ANNEX A.3

End of Scope



Banültepe
G. Banu MÜDERRİSOĞLU
 Secretary General